

Review Article

Innovations in Implant Dentistry: Exploring the Benefits of Short Implants

Guillermo Pardo-Zamora^{1*}, Antonio José Ortiz-Ruiz¹, Fabio Camacho-Alonso¹, José Francisco Martínez-Marco¹, Juan Manuel Molina-González¹, Núria Piqué-Clusella², Ascensión Vicente-Hernández¹

¹Department of General Dentistry and Implants, Faculty of Medicine and Dentistry, University of Murcia, 30008 Murcia, Spain.

²Microbiology Section, Department of Biology, Healthcare and Environment, Faculty of Pharmacy and Food Sciences, Universitat de Barcelona (UB), Av Joan XXIII, 27-31, 08028 Barcelona, Spain.

*E-mail ✉ gparza@um.es

Received: 01 February 2025; Revised: 28 April 2025; Accepted: 03 May 2025

ABSTRACT

The success of dental implants has led to an increased focus on methods to simplify the surgical procedure and increase implant longevity. With advancements in technology, the use of implants has expanded to include even the most complex cases. However, bone resorption following tooth extraction can lead to reduced bone volume and height, making standard implant placement challenging. In such situations, shorter implants have been suggested as an alternative, although their effectiveness as a treatment option remains unclear. Short implants offer advantages such as shorter treatment times, easier procedures, reduced patient morbidity, and lower costs. However, the literature also documents several biological complications, which could lead to their failure. This review examines the biomechanical factors, determinants of success, and the potential of short implants as a feasible solution for the rehabilitation of atrophic maxillary and mandibular alveolar ridges.

Keywords: Mandible, Dental implants, Short implants, Implantology, Maxilla

How to Cite This Article: Pardo-Zamora G, José Ortiz-Ruiz A, Camacho-Alonso F, Martínez-Marco JF, Molina-González JM, Piqué-Clusella N, et al. Innovations in Implant Dentistry: Exploring the Benefits of Short Implants. J Curr Res Oral Surg. 2025;5:33-43. <https://doi.org/10.51847/PHtKDGy2N7>

Introduction

Bone resorption following tooth extraction can significantly reduce both bone volume and height, making implant placement challenging [1]. After tooth loss, the posterior regions of the maxilla and mandible experience distinct patterns of bone loss. The maxilla typically shows slower vertical bone loss but experiences more significant horizontal loss in the buccal-palatal direction. This bone remodeling, along with maxillary sinus pneumatization, contributes to the vertical bone loss in 2 different directions. In contrast, vertical bone loss in the mandible mainly affects the vertical dimension, often reducing bone height but leaving a reasonable amount of horizontal bone. Planning for implant placement in atrophic posterior

arches is particularly complex due to these bone loss patterns and the proximity to important anatomical structures. The maxillary sinus expansion and the position of the mandibular canal (which is generally 10 mm or more above the inferior border of the mandible) typically result in reduced bone height in the posterior regions of both jaws [2]. To address this, techniques like guided bone regeneration (GBR), block grafts, sinus augmentation, and distraction osteogenesis have been proposed to restore lost bone height before implant placement. While these procedures have been successful, they come with increased complexity, sensitivity, and risks. As an alternative, short implants—defined as implants less than 10 mm—offer a less invasive solution for such cases. These implants simplify the surgical and prosthetic procedures,

making the treatment more accessible and cost-effective [1]. Key parameters to assess in such cases include marginal bone loss, implant survival rates, failure rates, and biological complications such as bleeding on probing and probing pocket depth [3]. Advances in implant surface geometry and texture have improved bone-to-implant contact, contributing to better primary stability during osseointegration. In cases of severely resorbed edentulous mandibles, six short implants may be used to support a fixed prosthesis, or 4 short implants may support an overdenture. In the edentulous maxilla, two additional short implants can be placed in the distal region alongside longer implants in the premaxilla to support a fixed prosthesis or overdenture [4]. This article aims to introduce the use of short implants in various clinical scenarios.

Results and Discussion

History

The development of the Bicon dental implant system in 1968 by Thomas Driskell marked a significant advancement in implant technology with the introduction of the 8-millimeter implant (**Figure 1**). Before this, the shortest conventional endosseous implants were 10 millimeters in length. Braunemark introduced the 7-millimeter implant, which led to the categorization of implants into two groups: short implants and ultra-short implants. “Short implants” are generally defined as those ranging from 7 millimeters to 10 millimeters in length, while “ultra-short implants” are those shorter than 7 millimeters. Additionally, the Bicon system introduced a 5-millimeter implant, which received approval from the Food and Drug Administration (FDA) in 2008 [5].

The Bicon system stands apart from other implant systems due to its unique design, largely influenced by Dr. Vincent J. Morgan, the founder and president of the company. Bicon believes that traditional threaded implants and high-speed drilling are unnecessary. Their system uses only three threads, which minimizes the risk of bone degeneration caused by heat and pressure, common issues with high-speed drilling. Driskell’s insight into the damage caused by fast drilling led him to adopt slow drilling methods in 1968. This approach offers numerous benefits, including greater patient comfort, the ability to harvest bone, improved visibility, and a significantly reduced risk of bone necrosis.



Figure 1. Bicon implant system

Short implants are highly effective because of the nature of the surrounding bone, which is primarily Haversian and cortical. These types of bone have superior mechanical properties compared to the appositional bone that surrounds traditional threaded implants. The effectiveness of an implant is largely influenced by its macro-geometry. For Bicon implants, the osteotomy is prepared using slow drilling at around 50 RPM, or slower with hand reamers. After the implant is inserted, the blood in its plateaus quickly transforms into Haversian bone, resembling cortical bone without osteoclastic activity. In contrast, when a threaded implant is inserted, pressure is applied to the bone, initiating osteoclastic activity [6].

While primary stability is an important factor, the first event following the insertion of a threaded implant is osteoclastic activity. The bone begins to regenerate as appositional bone, which lacks blood vessels, unlike the Haversian bone around short implants. This difference in bone characteristics may explain the higher success rate observed with short implants [7].

Thomas Driskell’s Bicon dental implant design, established over thirty years ago, has proven effective through numerous prosthetic restorations performed on this system [8].

Biomechanical considerations

Implant diameter

The bone crest receives the most stress during the implant process, while the apical section experiences significantly less stress. Increasing the length of an implant improves primary stability, but a wider diameter not only enhances primary stability but also increases the functional surface area at the bone crest, thereby improving the distribution of occlusal stresses. A finite element analysis by Himmlová *et al.* [9] demonstrated that a large implant diameter reduced stress around the implant neck and distributed masticatory forces more effectively. Similarly, Gavali

et al. [10] reported that increasing implant length enhances surface area and primary stability by increasing bone-implant contact (BIC). However, the functional surface area (FSA), which transmits compressive and tensile stresses to the bone, is primarily confined to the crestal 5 to 7 millimeters. Therefore, increasing the implant length beyond this point doesn't alter the FSA. On the other hand, a shorter implant with a wider diameter offers improved primary stability and a larger FSA [10].

Crown-to-implant ratio

A molar tooth can remain functional for decades with minimal root support, as demonstrated by ankylosed teeth. Advances in implant surface technology and load distribution have enabled the successful use of high crown-to-implant ratios. Meijer *et al.* reviewed studies and found that the crown-to-implant ratio for nonsplinted single-tooth implants ranged from 0.86 to 2.14, showing a low occurrence of biological or technical complications [11, 12]. According to Da Rocha Ferreira *et al.* [13], factors such as implant diameter, micro and macro geometry, implant-abutment connection, and bone quality and volume play a role in marginal bone stress. They concluded that the primary cause of marginal bone stress is excessive prosthetic height. This suggests a paradigm shift where prosthetic height is prioritized over implant length or crown-to-implant ratios. Further research is needed to develop new prosthetic designs that minimize stress at the marginal bone level (**Figure 2**) [13].




Figure 2. Crown-implant ratio

Bone quality

The success of short implants is largely dependent on the quality of the jawbone and its composition [14]. Research by Maló *et al.* [15] found that short implants in the mandible, which generally contains type I and II bone, exhibited a 99% success rate, whereas implants in the maxilla, predominantly made up of type III and IV bone, had a lower success rate of 92%. The porous structure of the maxillary bone is believed to have contributed to the varying success rates and implant

failures observed in these cases [15]. Galvão *et al.* [16] reviewed cases and noted that, regardless of the implant's surface treatment, failures were more frequent in areas with type III and IV bone, where bone density is lower. Short implants placed in regions with insufficient bone density may struggle with stability during both the initial placement and the healing period. Additionally, Liu *et al.* [17] found that in mandibles with inferior bone quality, surrounding bone in short implants was more prone to resorption. Tawil's patient series highlighted that the quality of bone plays a more significant role in determining implant longevity than the quantity of bone available (**Table 1**) [18].

Table 1. Bone density classification by Misch [19]

BONE		DENSITY
D1	> 1250 HU	Dense cortical bone
D2	850-1250 HU	Thick dense to porous cortical bone on crest and coarse trabecular bone
D3	350-850 HU	Thin porous cortical bone on the crest and fine trabecular bone within
D4	150-350 HU	Fine trabecular bone
D5	< 150 HU	Immature, non-mineralised bone

Absence of cantilevers

Cantilevers intensify the forces applied to the implant, with the force magnifying as the crown height increases. This creates multiple potential pivot points on the implant's body. By removing cantilevers, the predictability of treatment is improved, and biomechanics are enhanced [4, 19]. Thoma *et al.* [20] found that short implants without cantilevers produced radiological and comparable clinical outcomes to those with cantilevers after 5 years. However, the cantilevered implants tended to fail earlier, indicating they were subject to excessive force. Both options can be appropriate for clinical assessment [20].

Number of implants

Increasing the number of implants improves the distribution of occlusal stresses by expanding the surface area that the implants can handle [4].

Implant design

The thread design on the implant body is crucial for converting occlusal forces into favorable compressive loads at the bone interface. Threads are specifically designed to minimize micromovement that could impair osseointegration, maximize initial contact, improve primary stability, increase surface area, compress the bone, and help dissipate loads at the implant-bone interface [21]. Bolind *et al.* [22] reported that threaded implants showed greater bone contact,

while cylinder implants experienced more marginal bone loss.

Implant body designs with threads are effective in converting occlusal forces into compressive forces at the bone interface.

Factors affecting implant surface area

Increasing the number of threads per unit length on the axial plane increases the implant surface area in contact with the bone [4].

Thread depth

Deeper threads provide a larger contact surface area for the implant [4].

Thread shape (Figure 3)

Square threads provide a greater percentage of bone-to-implant contact when compared to V-shaped or reverse buttress threads [4]. V-shaped and reverse buttress threads produce ten times more shear force than square threads. The reduction of shear force at the thread-bone interface allows for better compressive load transfer, which is crucial in cases of low bone density, short implant lengths, or high force magnitude [23].

Implant surface treatment

The surface treatment of implants plays a critical role in the success of short implants. A rougher microtopography, as opposed to a smooth, turned surface, increases the bone-to-implant contact area and accelerates osseointegration. This is especially beneficial in compensating for an unfavorable crown-to-implant ratio [4].




Figure 3. a) V thread, b) reverse thread, and c) square thread

Surgical procedures

Two-stage surgical protocol

A two-step surgical protocol is recommended for short implants, as it ensures superior primary stability during the healing phase. Galvão *et al.* [16] suggested that this approach should be followed when placing short implants, with an interval of 4-6 months between the surgical and load stages for the maxilla, and 2-4 months for the mandible.

Modified surgical approach

To improve initial implant stability, modifications to the standard surgical process can be made, such as bypassing certain drilling steps, like the countersink or last drill, in the regular sequence. For better outcomes, narrow drills should be used for final bone preparation, especially in cases of poor bone quality, where soft bone drilling techniques should be employed [24].

Prosthetic considerations

Implant-abutment connection (Figure 4)

De Castro *et al.* [25] found that the Morse taper connection when compared to external hex abutment connections, resulted in less marginal bone loss and facilitated better bone growth around the implant shoulder.

Maeda *et al.* [26] conducted a study comparing stress distribution between external-hex and internal-hex connection systems using in vitro models. They observed that under horizontal loading, implants with external hex connections experienced greater strain at the cervical region, while internal hex implants demonstrated strain at the fixture tip. Additionally, internal hex connections provided a broader distribution of forces compared to the external hex design [26].




Figure 4. a) external hex, b) Morse taper, and c) internal hex

Platform switching

This technique helps preserve the crestal bone along the entire length of the implant, extending up to the collar [27-31].

Occlusal table

Reducing the size of the occlusal table can help minimize offset loads on the implant.

Incisal guidance

To accommodate the higher bite forces in the posterior regions, implants should adopt a biomechanical approach similar to that of natural teeth. Proper incisal guidance from the anterior teeth reduces lateral forces on the posterior teeth during mandibular movements [32-34].

Splinting

Splinting implants in soft bone enhances their functional surface area and minimizes the force transferred to the prosthesis, abutment screws, cement, and the bone-implant interface [4]. Ahumada-DeGirolamo *et al.* concluded that factors like implant size and its relation to crown height are significant considerations. Additionally, splinting adjacent crowns provides both biological and biomechanical benefits, especially for implants shorter than 8 mm or with an unfavorable crown-to-implant ratio. Each case must be evaluated individually, taking into account patient factors and parafunctional habits [35].

Factors affecting the survival rate of short implants

Various factors can increase the likelihood of implant failure, which are considered risk factors. These factors, as reported in the literature, include the bone quantity and quality, the age of the patient, the experience of the clinician, the site of implantation, implant length, axial stress, and maintenance of oral hygiene. Other significant risk factors for failure include chronic periodontitis, poor bone quality, systemic health issues, smoking, untreated infections or cavities, advanced age, improper implant placement, acentric loading, parafunctional habits, not enough implants, and the failure of the implant to properly integrate with both hard and soft tissues. Additionally, a poorly designed prosthesis can contribute to implant failure [36].

Smoking

Research on rats suggests that smoking has a more detrimental effect on cancellous bone than cortical bone. Bain's prospective study on implant surgery demonstrated that patients who quit smoking one week before and eight weeks after surgery experienced significantly fewer implant failures (11.8%) compared to those who continued smoking (38.5%). Smokers had slightly higher failure rates in compromised maxillary bone (2.6%) compared to nonsmokers (1.9%) [37, 38].

Implant location

While mandibular short implants generally have a higher success rate than maxillary ones, the location of short implants in the maxilla does not seem to be a significant risk factor for their success. Kim *et al.* found that implant failures were more common in the maxillary molar region, particularly with poor bone quality [39, 40]. However, Mezzomo *et al.*'s study [41] indicated a higher failure rate for single crowns supported by short implants in the mandible. The mandible, despite being cortical and rigid, is more prone to fractures, which may increase the risk of

implant failure compared to the maxilla [42]. Monje *et al.* [1] reviewed multiple cases and found no significant differences in the survival rates of short implants based on their anatomical location.

Periodontitis

Studies have shown that individuals with a history of periodontitis are at a higher risk for biological complications, for example, peri-implantitis and bone loss around the implant, leading to lower success rates and survival outcomes compared to those without such a background [43-45]. Peri-implantitis has been identified as a major contributor to the failure of short implants. In a study by Hasanoglu Erbasar *et al.* [39], it was determined that neither the length nor the diameter of the implant significantly influenced its success. However, having a history of periodontitis and smoking negatively impacted short implant success. Therefore, it is strongly recommended that patients with periodontitis undergo proper periodontal treatment before implant placement, along with a comprehensive periodontal care plan to ensure the longevity of short implants [46-49].

Splinting

The practice of splinting implants has shown positive effects, especially when shorter implants are connected to longer adjacent ones, which can contribute to their prolonged success [5]. According to Akça *et al.* [50], the group of splinted implants (97.7%) had slightly higher success rates compared to non-splinted implants (93.2%). While the failure rate for non-splinted implants was higher, particularly when shorter implants were used or in male patients, the success of splinted implants was not significantly affected by other factors. The increased failure rate in men could be linked to stronger bite forces, although this was not conclusively determined in the study.

Loading protocol

Weerapong *et al.* [51] found that immediate loading of mandibular molar implants showed similar survival rates, stability, and bone loss compared to traditional loading methods. Ayna *et al.* [52] observed favorable results with immediate loading protocols in the maxillary molar area, although the immediate loading group showed more bone loss and bleeding upon probing compared to delayed loading. This could be attributed to the poorer bone quality in the posterior maxilla or factors like the skill of the operator or the insertion torque [52]. In a study by Cannizzaro *et al.* [53], who applied a flapless approach for implant placement in the maxilla and mandible with immediate loading, the long-term results over nine years showed

which immediate placement of short implants was successful.

Short implants in the mandible

For patients with total edentulism in the posterior mandible, the lack of bone height due to the presence of the lower alveolar nerve can complicate implant placement. Techniques such as vertical ridge augmentation and lateralization of the alveolar nerve are employed to increase bone availability, though these procedures often carry risks, including graft contamination and infections post-surgery [54-73]. Sáenz-Ravello *et al.* [74] concluded that the use of short implants in atrophic mandibular ridges resulted in

fewer implant failures, less marginal bone loss, fewer biological issues, and higher patient satisfaction when compared to conventional implants placed after bone augmentation. Rosa *et al.* [75] also suggested that patients with mandibular atrophy may benefit from fixed full-arch restorations supported by short implants.

For patients with limited alveolar ridge height in the posterior mandible, there are three main treatment options: vertical ridge augmentation with implant placement, simultaneous implant placement with vertical ridge augmentation, and the use of short implants (**Figure 5a-5c**) [76].




Figure 5. a*) a possible treatment approach for the posterior mandible when the remaining ridge height is under 8 mm; b*) primary vertical bone augmentation should be carried out; c*) subsequently, standard-length implants should be placed; a**) a treatment approach for the posterior mandible when the remaining ridge height is between 8 and 10 mm; b**), c**) a short dental implant is advised; a***) a treatment approach for the posterior mandible when the vertical bone height exceeds 10 mm; b***, c***) standard-length implants are recommended, with primary vertical bone augmentation followed by implant placement.

Short implants in the maxilla

Sinus pneumatization, a natural physiological process, particularly affects the maxillary posterior region, accelerating bone loss. As a result, the quantity and

quality of bone often fall short of what is required for optimal three-dimensional (3D) implant placement [77]. To address the lack of sufficient bone height, procedures such as sinus augmentation using sinus

elevation or autogenous bone are employed to make room for the placement of standard implants. However, these sinus floor elevation surgeries can lead to complications such as upper lip numbness, sinus membrane tears, localized infections, swelling, hematomas, and maxillary sinusitis. Short implants emerged as a promising technique to assist in implant placement in cases of reduced alveolar bone, offering a way to avoid harm to critical structures [3].

Yan *et al.* [78] reviewed the use of short implants (≤ 6 mm) and found that they present a viable option for sinus floor elevation in patients with atrophic posterior maxillae, showing similar survival rates, reduced marginal bone loss (MBL), and fewer postoperative complications compared to traditional approaches. Several treatment alternatives are available, as shown in **Figure 6** [76].




Figure 6. a*) a possible treatment approach for the posterior maxilla when the vertical bone height exceeds 8 mm; b*, c*) for cases with a vertical ridge dimension greater than 8 mm, a transcresal sinus elevation method may be used if standard-length implants are chosen; a**) a treatment solution for the posterior maxilla with a vertical bone height ranging from 6–8 mm; b**, c**) in this case, the use of a short dental implant is recommended.

The choice of implant length plays a vital role in determining both the success of the prosthesis and the long-term survival of the implants. Placement in the posterior regions of the maxilla and mandible, where bone quality and volume are often insufficient, has traditionally been a challenging aspect of implantology. While longer implants can be utilized with advanced procedures such as bone augmentation and sinus lifts, these techniques can be more invasive. As a result, there is a growing preference for less invasive alternatives in areas with compromised bone. In this context, short implants are gaining recognition in the field of implant dentistry. Enhancements to the surface geometry and texture of implants have led to an expanded bone-implant contact area, which contributes to improved primary stability during the osseointegration process.

Conclusion

Short dental implants offer a viable solution for patients hesitant to undergo augmentation procedures.

When applied following strict clinical protocols, these implants can be deemed both safe and predictable, particularly for atrophic ridges in the maxilla and mandible. However, further high-quality research is necessary to assess the long-term success and effectiveness of short implants, considering factors such as implant location and overall success rates.

Acknowledgments: None

Conflict of Interest: None

Financial Support: None

Ethics Statement: None

References

1. Monje A, Fu JH, Chan HL, Suarez F, Galindo-Moreno P, Catena A, et al. Do implant length and width matter for short dental implants (<10 mm)?

- A meta-analysis of prospective studies. *J Periodontol.* 2013;84(12):1783-91.
2. Lum LB. A biomechanical rationale for the use of short implants. *J Oral Implantol.* 1991;17(2):126-31.
3. Esfahrood ZR, Ahmadi L, Karami E, Asghari S. Short dental implants in the posterior maxilla: a review of the literature. *J Korean Assoc Oral Maxillofac Surg.* 2017;43(2):70-6.
4. Jain N, Gulati M, Garg M, Pathak C. Short implants: new horizon in implant dentistry. *J Clin Diagn Res.* 2016;10(9):ZE14-7.
5. Shah AK. Short implants - When, where and how? *J Int Clin Dent Res Organ.* 2015;7(Suppl 1):132-7.
6. Sadeq ZA, Sabri LA, Al-Kinani KK. Natural polymer effect on gelation and rheology of ketotifen-loaded pH-sensitive in situ ocular gel (Carbapol). *J Adv Pharm Educ Res.* 2022;12(2):45-50.
7. Das N. The performance of short dental implants reaching a new height of precision. *Int J All Res Educ Sci Methods.* 2020;8(12):1234-43.
8. Morgan VJ, Isbaner G. The Bicon short implant. *Stomatol Edu J.* 2019;6(3):202.
9. Himmlová L, Dostálová T, Kácovský A, Konvicková S. Influence of implant length and diameter on stress distribution: a finite element analysis. *J Prosthet Dent.* 2004;91(1):20-5.
10. Gavali N, Chandak A, Waghmare P, Khadtare Y, Bhosale N, Verma H. Short dental implants: a systematic. *NeuroQuantology.* 2022;20(8):3649-60.
11. Birdi H, Schulte J, Kovacs A, Weed M, Chuang SK. Crown-to-implant ratios of short-length implants. *J Oral Implantol.* 2010;36(6):425-33.
12. Meijer HJA, Boven C, Delli K, Raghoobar GM. Is there an effect of crown-to-implant ratio on implant treatment outcomes? A systematic review. *Clin Oral Implants Res.* 2018;29(Suppl Suppl 18):243-52.
13. Da Rocha Ferreira JJ, Machado LFM, Oliveira JM, Ramos JCT. Effect of crown-to-implant ratio and crown height space on marginal bone stress: a finite element analysis. *Int J Implant Dent.* 2021;7(1):81.
14. Friberg B, Jemt T, Lekholm U. Early failures in 4,641 consecutively placed Brånemark dental implants: a study from stage 1 surgery to the connection of completed prostheses. *Int J Oral Maxillofac Implants.* 1991;6(2):142-6.
15. Maló P, de Araújo Nobre M, Rangert B. Short implants placed one-stage in maxillae and mandibles: a retrospective clinical study with 1 to 9 years of follow-up. *Clin Implant Dent Relat Res.* 2007;9(1):15-21.
16. Galvão FF, Almeida-Júnior AA, Faria-Júnior NB, Caldas SG, Reis JM, Margonar R. Predictability of short dental implants: a literature review. *RSBO (Online).* 2011;8(1):81-8.
17. Liu C, Xing Y, Li Y, Lin Y, Xu J, Wu D. Bone quality effect on short implants in the edentulous mandible: a finite element study. *BMC Oral Health.* 2022;22(1):139.
18. Tawil G, Younan R. Clinical evaluation of short, machined-surface implants followed for 12 to 92 months. *Int J Oral Maxillofac Implants.* 2003;18(6):894-901.
19. Misch CE. Contemporary implant dentistry. *Implant Dent.* 1999;8(1):90.
20. Thoma DS, Wolleb K, Schellenberg R, Strauss FJ, Hammerle CHF, Jung RE. Two short implants versus one short implant with a cantilever: 5-year results of a randomized clinical trial. *J Clin Periodontol.* 2021;48(11):1480-90.
21. Misch CE, Strong JT, Bidez MW. Scientific rationale for dental implant design. *Dent Implant Prosthet.* 2015;340-71. doi:10.1016/B978-0-323-07845-0.00015-4
22. Bolind PK, Johansson CB, Becker W, Langer L, Sevetz EB Jr, Albrektsson TO. A descriptive study on retrieved non-threaded and threaded implant designs. *Clin Oral Implants Res.* 2005;16(4):447-55.
23. Manikyamba YJ, Sajjan S, AV RR, Rao B, Nair CK. Implant thread designs: An overview. *Trends Prosthodont Dent Implantol.* 2018;8:11-20.
24. Fugazzotto PA, Beagle JR, Ganeles J, Jaffin R, Vlassis J, Kumar A. Success and failure rates of 9 mm or shorter implants in the replacement of missing maxillary molars when restored with individual crowns: preliminary results 0 to 84 months in function. A retrospective study. *J Periodontol.* 2004;75(2):327-32.
25. De Castro DS, de Araujo MA, Benfatti CA, de Araujo CD, Piattelli A, Perrotti V, et al. Comparative histological and histomorphometrical evaluation of marginal bone resorption around external hexagon and Morse cone implants: an experimental study in dogs. *Implant Dent.* 2014;23(3):270-6.
26. Maeda Y, Satoh T, Sogo M. In vitro differences of stress concentrations for internal and external hex implant-abutment connections: a short communication. *J Oral Rehabil.* 2006;33(1):75-8.
27. Vamvuka D, Teftiki A, Sfakiotakis S. Investigating the valorisation of refused derived

- fuel for energetic uses through its co-gasification with woody wastes. *World J Environ Biosci.* 2022;11(1):37-44.
28. Dirican S. A look at the change in water occupancy rates of Gölova Dam Lake, Turkey. *World J Environ Biosci.* 2022;11(1):34-6.
29. Alhazmi RA, Khayat SK, Albakri MH, Alruwaili WS, Bayazed HA, Almubarak SA, et al. An overview on the assessment and management of polycystic ovarian syndrome. *World J Environ Biosci.* 2022;11(1):17-23.
30. Alqurashi AM, Jawmin SA, Althobaiti TA, Aladwani MN, Almuebid AM, Alharbi JF, et al. An overview on nasal polyps' diagnosis and management approach. *World J Environ Biosci.* 2022;11(1):13-6.
31. Almuhanma MA, Alanazi MH, Al Ghamdi RN, Alwayli NS, Alghamdi IS, Qari AA, et al. Tachycardia evaluation and its management approach, literature review. *World J of Environ Biosci.* 2022;11(1):4-8.
32. Natarajan GP, Venkataraman SM, Pitchamuthu S, Rengaraj M. Impact of silicon seed priming on osmoregulators, antioxidants, and seedling growth of maize grown under chemo-stress. *World J Environ Biosci.* 2022;11(2):1-7.
33. Pimple NS. Virtual population analysis and recruitment pattern of *Osteobrama vigorsii* (Sykes, 1839) from Nira River, Bhor Maharashtra. *World J Environ Biosci.* 2022;11(1):53-9.
34. Chandra S, Meel RK. A systematic comparative study of *Morinda tinctoria* and *Vitex negundo* for their anti-ulcerogenic potential. *World J Environ Biosci.* 2022;11(1):45-52.
35. Ahumada-DeGirolamo D, Azocar A, Delpiano-Mesina C, Maldonado-Cortés P, Muñoz MA, Luque-Martínez I, et al. Splinting or non-splinting of fixed prostheses on adjacent implants: a critical review. *J Prosthodont Res.* 2024;68(2):206-14.
36. Porter JA, von Fraunhofer JA. Success or failure of dental implants? A literature review with treatment considerations. *Gen Dent.* 2005;53(6):423-46.
37. Bain CA. Smoking and implant failure--benefits of a smoking cessation protocol. *Int J Oral Maxillofac Implants.* 1996;11(6):756-9.
38. Bain CA, Moy PK. The association between the failure of dental implants and cigarette smoking. *Int J Oral Maxillofac Implants.* 1993;8(6):609-15.
39. Hasanoglu Erbasar GN, Hocaoglu TP, Erbasar RC. Risk factors associated with short dental implant success: a long-term retrospective evaluation of patients followed up for up to 9 years. *Braz Oral Res.* 2019;33:e030.
40. Kim YK, Hwang JY, Yun PY. Relationship between prognosis of dental implants and maxillary sinusitis associated with the sinus elevation procedure. *Int J Oral Maxillofac Implants.* 2013;28(1):178-83.
41. Mezzomo LA, Miller R, Triches D, Alonso F, Shinkai RS. Meta-analysis of single crowns supported by short (<10 mm) implants in the posterior region. *J Clin Periodontol.* 2014;41(2):191-213.
42. Villarinho EA, Triches DF, Alonso FR, Mezzomo LAM, Teixeira ER, Shinkai RSA. Risk factors for single crowns supported by short (6-mm) implants in the posterior region: a prospective clinical and radiographic study. *Clin Implant Dent Relat Res.* 2017;19(4):671-80.
43. Klokkevold PR, Han TJ. How do smoking, diabetes, and periodontitis affect outcomes of implant treatment? *Int J Oral Maxillofac Implants.* 2007;22 Suppl:173-202.
44. Pimentel SP, Shiota R, Cirano FR, Casarin RC, Pecorari VG, Casati MZ, et al. Occurrence of peri-implant diseases and risk indicators at the patient and implant levels: a multilevel cross-sectional study. *J Periodontol.* 2018;89(9):1091-100.
45. Mohanty R, Sudan PS, Dharamsi AM, Mokashi R, Misurya AL, Kaushal P. Risk assessment in long-term survival rates of dental implants: a prospective clinical study. *J Contemp Dent Pract.* 2018;19(5):587-90.
46. Patricia A, Hailemeskel B. Turmeric, black pepper, and lemon hot infusion for joint and musculoskeletal pain: a case report. *World J Environ Biosci.* 2024;13(1):36-8.
47. Doddapanen N, Lakshmegowda YK, Aardhya S, Rajashekar R, Doolgindachbaporn T, Nagaraju P. Environmental education, awareness and environmental ethics among pre-university students of Mysuru city, Karnataka, India. *World J Environ Biosci.* 2024;13(2):13-20.
48. Singar FA. Characterization of defatted cake prepared from egyptian olive's fruit (Wateeken Cultivar) and its biological activity. *World J Environ Biosci.* 2024;13(2):31-5.
49. Holachi H, Sathyanarayanaroa HT, Achar RR, Tulasidas VK. A sustainable way for integrated farming system: a case study on Bellary district of Karnataka, India. *World J Environ Biosci.* 2023;12(1):10-5.
50. Akça K, Iplikcioğlu H. Finite element stress analysis of the effect of short implant usage in

- place of cantilever extensions in mandibular posterior edentulism. *J Oral Rehabil.* 2002;29(4):350-6.
51. Weerapong K, Sirimongkolwattana S, Sastraruji T, Khongkhunthian P. Comparative study of immediate loading on short dental implants and conventional dental implants in the posterior mandible: a randomized clinical trial. *Int J Oral Maxillofac Implants.* 2019;34(1):141-9.
52. Ayna M, Wessing B, Gutwald R, Neff A, Ziebart T, Açil Y, et al. A 5-year prospective clinical trial on short implants (6 mm) for single tooth replacement in the posterior maxilla: immediate versus delayed loading. *Odontology.* 2019;107(2):244-53.
53. Cannizzaro G, Felice P, Leone M, Ferri V, Viola P, Esposito M. Immediate versus early loading of 6.5 mm-long flapless-placed single implants: a 4-year after loading report of a split-mouth randomised controlled trial. *Eur J Oral Implantol.* 2012;5(2):111-21.
54. Abdulqader AA, Rajab NA. Bioavailability study of posaconazole in rats after oral poloxamer P188 Nano-micelles and oral posaconazole pure drug. *J Adv Pharm Educ Res.* 2023;13(2):140-3.
55. Triyono T, Amijaya KA. Regular donor characteristics, inter-donation interval and the presence of subclinical anemia—A 3-year observational single-center study. *J Adv Pharm Educ Res.* 2023;13(2):118-23.
56. Arroyo-Fernández A, Blanco-Fernández MA, Lladó-Jordan G. Psychological therapy in overweight and obesity treatment centers. Do the centers include it on their websites? *J Adv Pharm Educ Res.* 2023;13(1):158-61.
57. Mustarichie R, Saptarini NM. Sirih (Piper betle) folium as new candidate for anti-herpes virus: in-silico study. *J Adv Pharm Educ Res.* 2023;13(1):46-50.
58. Nezhadrahim A, Shahri MM, Akbari NN. Effects of the Roux-en-Y gastric bypass on DM and renal function in obese patients. *J Adv Pharm Educ Res.* 2023;13(1):1-5.
59. Salih OS, Al-Akkam EJ. Pharmacokinetic parameters of ondansetron in rats after oral solution and transdermal invasomes gel: a comparison study. *J Adv Pharm Educ Res.* 2023;13(1):116-21.
60. Gofur NR, Gofur AR, Putri HM, Soesilaningtyas S, Halimah AN, Feriana D, et al. Drug choice to lowering risk contiguity with Morbus Hansen disease: a review article. *J Adv Pharm Educ Res.* 2023;13(1):73-9.
61. Feghhi M, Moghimi MA, Kasiri A, Farrahi F. Relative frequency of inherited retinal disorders in Khuzestan province, southern Iran. *J Adv Pharm Educ Res.* 2023;13(1):42-5.
62. Miranda LG, Amigo TR, de la Barra Ortiz HA. Validating an objective structured clinical examination to enhance assessment of clinical skills in physical therapy students. *J Adv Pharm Educ Res.* 2024;14(2):16-26.
63. Aldossary KM. Awareness and attitude towards premenstrual syndrome among foundation year students at Princess Nourah University: cross-sectional study. *J Adv Pharm Educ Res.* 2024;14(2):97-102.
64. Indratmoko S, Nurani LH, Wahyuningsih I. Enhancement of Icariin aphrodisiac effect by self nano emulsifying drug delivery system (SNEDDS) method. *J Adv Pharm Educ Res.* 2024;14(1):34-9.
65. Auwyang JA, Widiastih E. Can liver enzyme profile be a predictor of NAFLD in type-2 diabetes mellitus (T2DM) patients? *J Adv Pharm Educ Res.* 2024;14(1):8-12.
66. Mohammadi RG, Rajaei E, Ghanbaran A, Fatemi-Behbahani SA, Mowla K. Pregnancy and neonatal outcomes in pregnant patients with inflammatory arthropathy treated by etanercept; Retrospective study. *J Adv Pharm Educ Res.* 2024;14(2):58-62.
67. Madhukar CV. Production of potential bio-compost from household and market waste vegetables for the improvement of plant growth. *World J Environ Biosci.* 2022;11(2):15-9.
68. Kunder H, Gurusiddappa LH, Hanumanthappa BS, Kalikeri S. Antimicrobial activity and phytochemical analysis of solvent extraction of Citrus reticulata peels. *World J Environ Biosci.* 2022;11(2):20-5.
69. Almaghrabi SY, Hussein KS. Microbial profile of women with preterm premature rupture of membrane in Saudi Arabia: a retrospective study. *World J Environ Biosci.* 2022;11(2):43-7.
70. Al Blewi SM, Abuhamad RN, Abufara FA, Bedaiwi LA, Alotaibi JT, Alyami MA, et al. Epidemiology of epilepsy among pediatric patients in Tabuk city. *World J Environ Biosci.* 2022;11(2):48-53.
71. Shehata MA, AlDawsari AA, Saeedi AH, Ibrahim AlAyshan M, Saleh MA, Koshak HA, et al. Acute coronary syndrome diagnosis & management approach in emergency department: literature review. *World J Environ Biosci.* 2022;11(2):61-4.

72. Das SK, Karan S, Sen K. Biodiversity of avifauna in Chilkigarh, Jhargram, West Bengal, India. *World J Environ Biosci.* 2022;11(3):8-13.
73. Roy S, Laha I, Ray D, Choudhury L. Influence of climate change & environmental toxicants on epigenetic modifications. *World J Environ Biosci.* 2022;11(3):21-9.
74. Sáenz-Ravello G, Ossandón-Zúñiga B, Muñoz-Meza V, Mora-Ferraro D, Baeza M, Fan S, et al. Short implants compared to regular dental implants after bone augmentation in the atrophic posterior mandible: umbrella review and meta-analysis of success outcomes. *Int J Implant Dent.* 2023;9(1):18.
75. Rosa A, Pujia AM, Arcuri C. Complete full arch supported by short implant (< 8 mm) in edentulous jaw: a systematic review. *Appl Sci.* 2023;13(12):7162.
76. Thoma DS, Cha JK, Jung UW. Treatment concepts for the posterior maxilla and mandible: short implants versus long implants in augmented bone. *J Periodontal Implant Sci.* 2017;47(1):2-12.
77. Lozano-Carrascal N, Anglada-Bosqued A, Salomó-Coll O, Hernández-Alfaro F, Wang HL, Gargallo-Albiol J. Short implants (<8mm) versus longer implants ($\geq$ 8mm) with lateral sinus floor augmentation in posterior atrophic maxilla: a meta-analysis of RCT's in humans. *Med Oral Patol Oral Cir Bucal.* 2020;25(2):e168-79.
78. Yan Q, Wu X, Su M, Hua F, Shi B. Short implants ($\leq$ 6 mm) versus longer implants with sinus floor elevation in atrophic posterior maxilla: a systematic review and meta-analysis. *BMJ Open.* 2019;9(10):e029826.